

Reagents needed for each of the PoolLab 1.0 photometer's parameters

parameter	range	reference reagent	description
Active Oxygen	0.0 - 30.0 mg/l	PRLR-RD4-050	Reagent DPD N°4, 50 tablets
Alkalinity	0 - 300 mg/l	PRLR-005-050	Alkalinity - M, 50 tablets
Bromine	0.0 - 13.5 mg/l	PRLR-RD1-050	Reagent DPD N°1, 50 tablets
		PRLR-HGC-050	Reagent Glycine, 50 tablets
Chlorine	0.00 - 6.00 mg/l	PRLR-RD1-050	Reagent DPD N°1, 50 tablets
		PRLR-RD3-050	Reagent DPD N°3, 50 tablets
Chlorine Dioxide	0.0 - 11.4 mg/l	PRLR-RD1-050	Reagent DPD N°1, 50 tablets
		PRLR-HGC-050	Reagent Glycine, 50 tablets
Cyanuric Acid	0 - 160 mg/l	PRLR-020-050	Cyanuric acid, 50 tablets
Calcium Hardness	0 - 500 mg/l	POLR-CH1-020	Calcium Hardness N°1 (20 ml, 50 tests)
		POLR-CH2-020	Calcium Hardness N°2 (20 ml, 50 tests)
Total Hardness	0 - 500 mg/l	PRLR-481-050	Total Hardness N°1 (20 ml, 50 tests)
		PRLR-482-050	Total Hardness N°2 (10 ml, 50 tests)
Hydrogen Peroxide LR	0.00 - 2.90 mg/l	PRLR-024-050	Hydrogen peroxide LR, 50 tablets
Hydrogen Peroxide HR	0 - 200 mg/l	POLR-APT-050	Acidifying reagent, 50 tablets
		POLR-HPH-050	Hydrogen Peroxide HR reagent, 50 tablets
Ozone	0.00 - 4.00 mg/l	PRLR-RD1-050	Reagent DPD N°1, 50 tablets
		PRLR-RD3-050	Reagent DPD N°3, 50 tablets
		PRLR-HGC-050	Reagent Glycine, 50 tablets
pH	6.5 - 8.4 pH	PRLR-038-050	pH MR Phenol red, 50 tablets